ITEM NO	SRM CODE	ITEM DESCRIPTION	GROUP
MMG110044	4226160000100	SCALE AUTOPSY TABLE TOP	MMG110044
MMG110044-D1	400000054300	MAIN AREA DELIVERY COST RELATED TO SCALE AUTOPSY TABLE TOP	MMG110044
MMG110044-D2	400000054400	REMOTE AREA DELIVERY COST RELATED TO SCALE AUTOPSY TABLE TOP	MMG110044
MMG110044-W3	4200011499000	3rd YEAR WARRANTY COST RELATED TOSCALE AUTOPSY TABLE TOP	MMG110044
MMG110044-W4		4th YEAR WARRANTY COST RELATED TOSCALE AUTOPSY TABLE TOP	MMG110044
MMG110044-W5		5th YEAR WARRANTY COST RELATED TOSCALE AUTOPSY TABLE TOP	MMG110044
NMG10001-001		DAYLIGHT LAMP RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-D1		MAIN AREA DELIVERY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-D2		REMOTE AREA DELIVERY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P001		PREPARING THE SITE TO START WORK BY REMOVING. DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILD	
NMG10001-P002		SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P003		PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER	
NMG10001-P004		SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P004		SUPPLY AND INSTALLATION OF CABLE THAT RECATED TO AIR INLET WITH INTEGRALED DATEIGHT. SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO AIR INLET WITH INTEGR	
NMG10001-P006		CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA. RELATED TO AIR INLET WITH INTEGRAL INC. OUT TO AIR I	NMG10001
NMG10001-P007		DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO AIR INLET WITH INTEGRATED DAYLIG	
NMG10001-P007		REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO AIR INLEG WITH INTEGRATED DAYLIGHT	
NMG10001-P008		' ' '	NMG10001
		REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO AIR INLET WITH INTEG	
NMG10001-P010	4200011165600	CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO AIR IN	
NMG10001-P011		SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL RELATED TO AIR INLET WITH	
NMG10001-P012		SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO AIR INLET WIT	
NMG10001-P013		SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE	
NMG10001-P014		SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY C	
NMG10001-P015		SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F	
NMG10001-P016		SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P017		SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P018		SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P019		SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P020		SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P021		SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P022		SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P023		SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREA	
NMG10001-P024		INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO AIR INLET WITH INTEGRATE	NMG10001
NMG10001-P025	4200011167100	INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P026	4200011167200	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P027		SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P028	4200011167400	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P029	4200011167500	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P030	4200011167600	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P031	4200011167700	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P032	4200011167800	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P033	4200011167900	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P034	4200011168000	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P035	4200011168100	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P036	4200011168200	SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW	NMG10001
NMG10001-P037	4200011168300	SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P038	4200011168400	SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P039	4200011168500	SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH R.145. AS PER HOSPITAL STANDARDS. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P040		SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO AIR INLE	
NMG10001-P041		SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO AIR INLET W	
NMG10001-P042		RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPES,	
NMG10001-P043	4200011168900	FIRE WATER LINE RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P044	4200011169000	SPRINGKLER HEADS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001
NMG10001-P045		DRAIN PIPES RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT	NMG10001

NMG10001-P047 4200011169300 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P048 4200011169400 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P050 4200011169500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P051 4200011169600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P051 4200011169700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P052 4200011169900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLIATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NMG10001-P055 4200011170100 SUPPLY AND INSTALLIANG PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR INMG10001-P056 4200011170100 SUPPLY AND INSTALLIANG AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 4200011170100 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 3rd YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT	MG10001 MG10001 MG10001 MG10001
NMG10001-P048 4200011169400 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P049 4200011169500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P050 4200011169600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P051 4200011169700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P052 4200011169800 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P054 420001117000 BE SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P054 420001117000 BE SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P055 420001117000 BE SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 420001115000 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 420001115000 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR THAT THE NEAR STANDARD TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 420001165100 SUPPLY AND INSTALLATION OF A/C BUPPLY AND INSTALLATION OF A/C BU	1G10001 1G10001 1G10001 1G10001 1G10001 1G10001 1G10001 1G10001 1G10001
NMG10001-P049 4200011169500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P050 4200011169600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P051 4200011169700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO AIR INLE NMG NMG10001-P052 4200011169800 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P053 4200011169900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P055 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 4200011170200 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 ST YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 Sth YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W6 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W6 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W6 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W7 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W6 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W7 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W6 420001165100 AIR INLET WITH INTEGRATED DAYLIGHT	MG10001 MG10001 MG10001 MG10001 MG10001 MG10001 MG10001 MG10001
NMG10001-P050 4200011169600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P051 4200011169700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P052 4200011169800 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P053 4200011169900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NING NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLCHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P054 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4 HYEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 4000000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	MG10001 MG10001 MG10001 MG10001 MG10001 MG10001
NMG10001-P051 4200011169700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO AIR INLE NMG NMG10001-P052 4200011169800 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P053 4200011169900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NMG NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR NMG10001-P055 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011501300 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 4000000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	MG10001 MG10001 MG10001 MG10001 MG10001
NMG10001-P052 4200011169800 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P053 4200011169900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NMG NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NMG NMG10001-P055 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 4000000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	MG10001 MG10001 MG10001 MG10001
NMG10001-P053 4200011169900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO AIR INLET NMG NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FO NMG NMG10001-P055 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 4200011170200 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 3rd YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-D1 4000000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	1G10001 1G10001 1G10001
NMG10001-P054 4200011170000 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR NMG10001-P055 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 4200011170200 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 3rd YEAR WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	1G10001 1G10001 1G10001
NMG10001-P055 4200011170100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-P056 4200011170200 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 3rd YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT NMG10001-P056 4200011651500 SUPPLY AND INSTALL CHILLER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY AND INSTALL CHILLER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY AND INSTALL CHILLER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY AND INSTALL CHILLER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY AND INSTALL CHILLER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651300 SIPPLY HOT & WARRANTY COST RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT N	1G10001 1G10001
NMG10001-P056 4200011170200 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO AIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W3 4200011651300 3rd YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT NMG	1G10001
NMG10001-W3 4200011651300 3rd YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT NMG	
NMG10001-W4 4200011651400 4th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001-W5 4200011651500 5th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHT NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT NMG	astooni l
NMG10001-W542000116515005th YEAR WARRANTY COST RELATED TOAIR INLET WITH INTEGRATED DAYLIGHTNMGNMG100014111250000200AIR INLET WITH INTEGRATED DAYLIGHTNMGNMG10003-D1400000054300MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHTNMG	1G10001 1G10001
NMG10001 4111250000200 AIR INLET WITH INTEGRATED DAYLIGHT NMG10003-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT NMG	1G10001 1G10001
NMG10003-D1 4000000054300 MAIN AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	
NMG10003-DZ 4000000054400 REMOTE AREA DELIVERY COST RELATED TO CABINET DRYING WITH UV LIGHT	
AND ADDRESS OF THE PROPERTY OF	
	1G10003
	1G10003
	1G10003
	1G10003
	1G10004
NMG10004-D2 400000054400 REMOTE AREA DELIVERY COST RELATED TO CABINET EVIDENCE	1G10004
NMG10004-W3 4200011652100 3rd YEAR WARRANTY COST RELATED TOCABINET EVIDENCE NMG	1G10004
NMG10004-W4 4200011652200 4th YEAR WARRANTY COST RELATED TOCABINET EVIDENCE NMG	1G10004
NMG10004-W5 4200011652300 5th YEAR WARRANTY COST RELATED TOCABINET EVIDENCE NMG	1G10004
NMG10004 4219190502000 CABINET EVIDENCE NMC	1G10004
NMG10006-001 4219191400300 WALL MOUNTED REFRIGERATOR RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-002 4219191400400 WALL MOUNTED WARMER RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-D2 400000054400 REMOTE AREA DELIVERY COST RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-P001 4200011170300 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILD INMG	IG10006
NMG10006-P002 4200011170400 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-P003 4200011170500 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER NMG	IG10006
	1G10006
NMG10006-P005 4200011170700 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO CABINET WALL MOUNTING	
	1G10006
NMG10006-P007 4200011170900 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO CABINET WALL MOUNTED INSTRUM NMG	
	1G10006
NMG10006-P009 4200011171100 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO CABINET WALL MOUNTING	
NMG10006-P010 4200011171200 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO CABIN MMG	
NMG10006-P011 4200011171300 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO CABINET WALLIAMS	
NMG10006-P012 4200011171400 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO CABINET WALLING.	
NMG10006-P013 4200011171500 SUPPLY AND INSTALLATION OF MICHORISED SIGNING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATERIAL AS AFRICADOR FRAME. SIZE AS PER DRAWING. SIZE AS	
NMG10006-P013 4200011171500 SUPPLY AND INSTALLATION OF SINGLE SWING BOOKS, INCLUDING REACT BOTT HARDWARE HINGES AND BOOK HANDLES ON HOLLOW METAL BOOK FRANKE. SIZE AS PER DRAWING. MATERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY ON MICE.	
NMG10006-P015 420001171700 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER FINMS	
	1G10006
	1G10006
	1G10006
NMG10006-P019 4200011172100 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT	1G10006
	1G10006
NMG10006-P020 4200011172200 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-P020 4200011172200 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P021 4200011172300 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006
NMG10006-P020 4200011172200 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P021 4200011172300 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P022 4200011172400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT	
NMG10006-P020 4200011172200 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG10006-P021 4200011172300 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT NMG	1G10006

NMG10006-P025	4200011172700	INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P026	4200011172800	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P027	4200011172900	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P028	4200011173000	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P029	4200011173100	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P030	4200011173200	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P031	4200011173300	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P032	4200011173400	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P033		SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P034	4200011173600	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P035	4200011173700	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P036		SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW	
NMG10006-P037	4200011173900	SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P038	4200011174000	· ·	NMG10006
NMG10006-P039	4200011174100		NMG10006
NMG10006-P040	4200011174100	SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO CABINITY	
NMG10006-P041	4200011174200		
NMG10006-P041	4200011174300	RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPES	
NMG10006-P042			NMG10006
NMG10006-P043	4200011174500	SPRINGKLER HEADS RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P044	4200011174700	DRAIN PIPES RELATED TO CABINET WALL MOUNTED INSTRUMENT	
NMG10006-P045		SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006 NMG10006
NMG10006-P047	4200011174800	SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P047	4200011174900	SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO CABINET WALL MOUNTED INSTRUMENT	
NMG10006-P048		SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006 NMG10006
NMG10006-P049	4200011175100	SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P050	4200011175200	CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO CABIN	
NMG10006-P051		SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P052	4200011175400	SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO CABINE	
NMG10006-P053	4200011175600	BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE STSTEM WHICH INCLUDE PIPING WORS, TAPPED TO THE NEAREST BOILDING OTHER TSTSTEM. RELATED TO CABINE	_
NMG10006-P055		SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-P055	4200011175700		_
NMG10006-P036	4200011173800	SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO CABINET WALL MOUNTED INSTRUMENT 3rd YEAR WARRANTY COST RELATED TOCABINET WALL MOUNTED INSTRUMENT	NMG10006
NMG10006-W3	4200011652500		NMG10006
NMG10006-W4	4200011652600	Sth YEAR WARRANTY COST RELATED TOCABINET WALL MOUNTED INSTRUMENT	NMG10006 NMG10006
	4219190500900	CABINET WALL MOUNTED INSTRUMENT	
NMG10006			NMG10006
NMG10008-001	4226180601800 4226180601900	CENTRAL STATION CONTROL RELATED TO FREEZER 03 BODY RECORDER RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-002	4000000054300	MAIN AREA DELIVERY COST RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-D1	†		NMG10008
NMG10008-D2 NMG10008-P001	4000000054400	REMOTE AREA DELIVERY COST RELATED TO FREEZER 03 BODY	NMG10008
	4200011175900	PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILD BY AND INSTALLATION OF STAINLIESS STEEL DRAINGE CRILL WIDTH NOT LESS THAN 12 CM BELATED TO EREFTED 03 PODY.	
NMG10008-P002		SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P003	4200011176100	PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PE	
NMG10008-P004	 	SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P005	4200011176300	SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P006	4200011176400	CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P007		DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P008	4200011176600	REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P009	4200011176700	REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL) INTERNAL), WINDOW, DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P010	4200011176800		_
NMG10008-P011	4200011176900	SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER 03 B	
NMG10008-P012	4200011177000	SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER 03	_
NMG10008-P013	4200011177100	,	
NMG10008-P014	4200011177200	SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY	
NMG10008-P015	42000111//300	SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER	FINING10008

	T		1
NMG10008-P016		SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P017	+	SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P018	4200011177600	SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P019	4200011177700	· ·	NMG10008
NMG10008-P020	4200011177800		NMG10008
NMG10008-P021	4200011177900	SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P022	4200011178000	SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P023		SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BRE	
NMG10008-P024	4200011178200	INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P025	4200011178300	INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P026		SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P027	4200011178500	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P028	4200011178600	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P029	<u> </u>	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P030	4200011178800	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P031	4200011178900	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P032	+	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P033	4200011179100	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P034	4200011179200	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P035	+	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P036	4200011179400	SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW	
NMG10008-P037	4200011179500	SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P038	4200011179600	SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P039	4200011179700	SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RJ45. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P040	4200011179800	SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZI	
NMG10008-P041	4200011179900	SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 03	3 NMG10008
NMG10008-P042	4200011180000	RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPE	S, NMG10008
NMG10008-P043	4200011180100	FIRE WATER LINE RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P044	+	SPRINGKLER HEADS RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P045	4200011180300	DRAIN PIPES RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P046	+	SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P047	4200011180500	SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P048	4200011180600	SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P049	4200011180700	SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P050	4200011180800	SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P051	4200011180900	CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZ	EFNMG10008
NMG10008-P052	4200011181000	SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P053	4200011181100	SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZE	R NMG10008
NMG10008-P054	4200011181200	BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR	OFNMG10008
NMG10008-P055	4200011181300	SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-P056	4200011181400	SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 03 BODY	NMG10008
NMG10008-W3	4200011652900	3rd YEAR WARRANTY COST RELATED TOFREEZER 03 BODY	NMG10008
NMG10008-W4	4200011653000	4th YEAR WARRANTY COST RELATED TOFREEZER 03 BODY	NMG10008
NMG10008-W5	4200011653100	5th YEAR WARRANTY COST RELATED TOFREEZER 03 BODY	NMG10008
NMG10008	4226180600000	FREEZER 03 BODY	NMG10008
NMG10009-001	4226180602000	CENTRAL STATION CONTROL RELATED TO FREEZER 06 BODY	NMG10009
NMG10009-002	4226180602100	RECORDER RELATED TO FREEZER 06 BODY	NMG10009
NMG10009-D1	400000054300	MAIN AREA DELIVERY COST RELATED TO FREEZER 06 BODY	NMG10009
NMG10009-D2	400000054400	REMOTE AREA DELIVERY COST RELATED TO FREEZER 06 BODY	NMG10009
NMG10009-P001	4200011181500	PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUIL	D NMG10009
NMG10009-P002	4200011181600	SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO FREEZER 06 BODY	NMG10009
NMG10009-P003	4200011181700	PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PE	ER NMG10009
NMG10009-P003 NMG10009-P004	4200011181700 4200011181800		NMG10009
	4200011181800		
NMG10009-P004	4200011181800 4200011181900	SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO FREEZER 06 BODY	NMG10009

MINISTEDIO POSO PROPERTY PROVIDED OF PERSYMEN WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT RELATED TO PERSEZE BEODY MINISTEDIO PROVIDED OF	NMG10009-P007	4200011182100	DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO FREEZER 06 BODY	NMG10009
INACIDIDATE POR SEQUELLISATION OF SECURITY OF WALLS CETTAINED TO FREEZE OF SECURITY OF WALLS CETTAINED TO FREEZE OF SECURITY OF WALLS CETTAINED FOR SECURITY O		-	· · · · · · · · · · · · · · · · · · ·	
NASCOGOPPO 20001118200 CONTRICTION OF MASONNY WALLS WITH RER RATED AND MOSCINER RESCRIAT, FIRE OR BOCKWOOD SURDEPROPOR INCLINION OF MOTORIZED SWINN OF MOTORIZED AUGMOND ORDS. REQUIRED ORDS. AUGMOND				
NAMISCOOP 1011. 40001118700. SUPPLY ARD INSTALLATION OF MOTORIZED SOUND GOOR, INCLIDIONS COVER-RAD MICHANISMS, SEE A SPE DAWNER, MATERIAL AS APPROVED BY HOSPITAL, REATION OF HEEZER OR GOVERNOODED. NAMISCOOP 1012. 40001118700. SUPPLY ARD INSTALLATION OF SHINGL SWING GOORS, INCLIDIONS CHAPTER OF THE CONTROL STATE OF THE CONTROL SHIP AND INSTALLATION OF SHINGL SWING GOORS, INCLIDIONS CHAPTER OF THE CONTROL SHIP AND INSTALLATION OF SHINGL SWING GOORS, INCLIDIONS CHAPTER OF THE CONTROL SWING				
INACIDIOS POLIZ ADDITISTATION SUPPLY AND INSTALLATION OF MOTORIZO SUDNING DOOR, INCLUDING CHAPTER AND ADDITISTATION OF MOTORIZO SERVICE SHIP SUBJECT SERVICES AND ADDITISTATION OF MOTORIZON SERVICES AND ADDITISTATION OF ADDITISTATION		+		_
INASCISIOS POTOS 200011182/00 300001182/00 3000001182/00 300000182/00 300000182/00 300000182/00				
MISSIOSOP-1914 MOSIOSOP-1915 MOSIONITIES DISTRICT FOR WAS HERE FOR STANLE FINDERS SENT CLOSS WALL ART BACTERIA PAINT, INCLUDING SURFACE PREPARATION WITH A NIMIMUM OF THREE COATS AND PREFERANCE MISSIONATION OF HEAVY-DUTY THREE MAINS SENT CONTROL PREAD PAINT AND STANLE FOR HEAVY PAINT STANLE F			' '	
MIGISIOR-1913 MORISIOR-1913 MORISIOR-1914 MORISIOR-1915 MORISIOR-1917 MORISI	-	+		
MIGHIOROPHOSE DODOTISATION SUPERVI AND INSTALL RESIDENCE ALUNDAY OF RESERVE OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL SECRETAL HUNG DODO OR SUDING DOOR SELENCE TO THE REZER OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL SECRETAL HUNG DOOR OR SUDING DOOR SELENCE TO THE REZER OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL SECRETAL SECRETAL BAND THE RECEIVED TO THE REZER OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL COMPRESSOR SUBBATION RESISTANCE BASK AS PER ROSPITAL REQUIREMENTS RELATED TO THE REZER OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL COMPRESSOR SUBBATION RESISTANCE BASK AS PER ROSPITAL REQUIREMENTS RELATED TO THE REZER OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL COMPRESSOR SUBBATION RESISTANCE BASK AS PER ROSPITAL REQUIREMENTS RELATED TO THE REZER OF BODY MIGHIOROPHOSE ADDOTISATION SUPERVI AND INSTALL COMPRESSOR SUPERVI AND INSTALLATION OF ANY EXCHANGE WITH ALL RESISTANCE COMPUTER WITH SUPERVI AND INSTALLATION OF ANY EXCHANGE WITH ALL RESISTANCE COMPUTER AND INSTALLATION OF ANY EXCHANGE WITH ALL RESISTANCE COMPUTER AND INSTALLATION OF THE REQUIRED EXCEPTIVE COMPUTER AND INSTALLATION				
MAG 10009 PRIT MOR 10009				+
MMSI0009-P013 MOSIO00183800 SUPPLY AND INSTALL SINCE WOODEN DOOR ANTIBACTIFICAL & ANTIFIER SELECT OF RECEZER OS BODY NMSI0009-P014 MOSIO00183800 SUPPLY AND INSTALL COMPRESSOR PARKINGS AND INSTALLATION P				
INAGEGIODP-POST 2 10001118303 SUPPY AND INSTALL STAINLESS STEEL WALL PANELING ABOUND EQUIPMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 10001118305 SUPPY AND INSTALL COMMESSON SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 1000118305 SUPPY AND INSTALL COMMESSON SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 1000118305 SUPPY AND INSTALL COMMESSON SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 1000118305 SUPPY AND INSTALL COMMESSON CARE AS PER HOSPITAL REQUIREMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 1000118305 SUPPY AND INSTALL COMMESSON CARE AS PER HOSPITAL REQUIREMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 1000118305 SUPPY AND INSTALL COMMESSON CARE AS PER HOSPITAL REQUIREMENTS RELATED TO PREZER OB BODY NAGIGOODP-POST 2 1000118305 SUPPY AND INSTALLATION OF A THE REQUIRED STREAM COMMESSON CARE AS PER HOSPITAL REQUIRED STREAM COMMESSON CARE AS PER HOSPIT		-		
NAMISTOROPHOZO ACCOUNTAGE ON THE PROPERTY AND INSTALL COMPRESSOR WIRBATION ESSISTANCE BASE A SPER HOSPITAL REQUIREMENTS RELATED TO FREEZER OS BODY NAMISTOROPHOZO ACCOUNTAGE ON THE PROPERTY AND INSTALL COMPRESSOR WIRBATION OF EVER A PRE NOSTITAL REQUIREMENTS RELATED TO FREEZER OS BODY NAMISTOROPHOZO ACCOUNTAGE ON THE PROPERTY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER OS BODY NAMISTOROPHOZO ACCOUNTAGE ON THE PROPERTY AND INSTALL POR POWER DISTRIBUTION BOARD IP FREAT PROPERTY AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FREAT PROPERTY AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FREAT PROPERTY AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FREAT PROPERTY AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FRANCE AND VIEW AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FRANCE AND VIEW AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FRANCE AND VIEW AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FRANCE AND VIEW AND VIEW AND INSTALLAD ON POWER POWER AND INSTALLAD ON POWER DISTRIBUTION BOARD IP FRANCE AND VIEW AND VIEW AND INSTALLAD ON POWER POWER AND INSTALLAD ON POWER AND				
INGGIOGOP-P022 (20001118800) SUPPLY AND INSTALL COMPRESSOR CASE A PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 08 BODY MINGGIOOSP-P023 (20001118800) SUPPLY AND INSTALL OMPRESSOR CASE A PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 08 BODY MINGGIOOSP-P024 (20001118800) SUPPLY AND INSTALL DED (POWER DISTRIBUTION BOARD) PER FINAL/FLECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CRICUIT BREA MINGGIOOSP MINGGIOOSP-P025 (20001118800) SUPPLY AND INSTALLATION OF FOR BECKER OS ON X SOCIAL WITHOUT WITH WITHOUT AND WITHOUT SUPPLY AND INSTALLATION OF PROVER SOCIAL PROVINCE AND SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 35 MMAC CABLE WIRES RELATED TO FREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 35 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 35 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 35 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 35 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 35 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 30 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 30 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PC POWER CABLES, 3 PHASE, 380V, 30 MMAC CABLE WIRES RELATED TO TREEZER OB BODY MINGGIOOSP-P027 (20001118910) SUPPLY AND INSTALLATION OF THE REQUIRE			·	
INGGIOGO-PROZ ACCOUNTINGSOOD STATES AND STALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZE OS BODY MINGGIOGO-PROZ ACCOUNTINGSOOD STATES AND STALLATION OF ANY BACTERIAL LED LIGHTING FRANCE PER HAVELECTRICAL DORAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREAK MINGGIOCOPA PROX ACCESSORIES, CIRCUIT BREAKER, SOLOW ACC				
NMG10009-P023 4200011138300 INSTALLATION OF THE REQUIRED XEPT VAN DISTALLATION OF THE XEPT VA				
MISSIGNOSP-9024 A200011138300 INSTALLATION OF ANTI BACTERIAL ED LIGHTING FIXTURES (50 CM X DC CM) WITH SWITCHES AS PER HOSPITAL STANDARDS, RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CARLES; 3 PHASE, 380V, 3 S MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 3 S MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 5 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 5 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 7 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 10 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 10 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 10 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF MAIN DESCONACT SWITCH IN THE COURT FOR THE REQUIRED XILPEPINC POWER CABLES; 3 PHASE, 380V, 10 MMZ CABLE WIRES RELATED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF MAIN DESCONACT SWITCH IN THE REQUIRED TO FREEZER 06 BODY NNG510009- NNG510009- P026 A200011138000 SUPPLY AND INSTALLATION OF MAIN DESCONACT SWITCH IN THE REQUIRED TO FREEZER 06 BODY NN			·	
MMG10009-P025 A200011138300 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC				
NMG10099-9236 4200011184000 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 35 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9208 4200011184200 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 50 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9208 4200011184200 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 50 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9304 420001118400 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 70 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9304 420001118400 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 90 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9324 420001118400 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 10 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9324 420001118400 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE, 380X, 10 MM2 CABLE WIRES RELATED TO RECEZER OB BODY MMG10009 NMG10009-9324 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNICT SWITCH IN THE EQUIRMENT ROOM, 20 AMP RELATED TO RECEZER OB BODY MMG10009 NMG10009-9324 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNICT SWITCH IN THE EQUIRMENT ROOM, 20 AMP RELATED TO RECEZER OB BODY MMG10009 NMG10009-9324 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNICT SWITCH IN THE EQUIRMENT ROOM, 20 AMP RELATED TO RECEZER OB BODY NMG10009 NMG10009-9324 420001118500 SUPPLY AND INSTALLATION OF FIAL PROPERTY OF THE POWER PART OF THE POWER PART OF THE POWER PART OF THE				
NMG1009-P027 (2000118410D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009 NMG10009-P029 (2000118430D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009 NMG10009-P029 (2000118430D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009 NMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES, 3 PHASE, 38DV, SO MMZ CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED 10 FREEZER OB BODY MMG10009-P031 (2000118450D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED 10 FREEZER OB BODY MMG10009-P031 (200011850D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED 10 FREEZER OB BODY MMG10009-P031 (200011850D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT SREAKER, SLOW BUNNING MAIN CABLE WIRES RELATED 10 FREEZER OB BODY MMG10009-P032 (200011850D) SUPPLY AND INSTALLATION OF MAIN DISCONNECT				
NINGGIOODP-PO28 4200011184200 SUPPLY AND INSTALLATION OF THE REQUIRED XPE/PVC POWER CABLES, 3 PHASE. 380W, 70 MMZ CABLE WIRES RELATED TO REEZER 06 BODY MINGGIOODP MI				
NINGS1000P-P029 ADDIDITISATION SUPPLY AND INSTALLATION OF THE REQUIRED XIPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MMZ CABLE WRIES RELATED TO FREEZER 06 BODY NINGS1000P NINGS1000P P031 ADDIDITISATION ADDITISATION OF THE REQUIRED XIPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MMZ CABLE WRIES RELATED TO FREEZER 06 BODY NINGS1000P NINGS1000P P032 ADDITISATION OF THE REQUIRED XIPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MMZ CABLE WRIES RELATED TO FREEZER 06 BODY NINGS1000P NINGS100P NINGS				
NUMSIGOODP-PO30 ADDITISEADO SUPPLY AND INSTALLATION OF THE REQUIRED XUPE/PVC POWER CABLES; 3 PHASE 380V, 120 MM2 CABLE WIRES RELATED TO FREEZER 06 BODY NUMSIGOODP NUMSI				
NMG1009-P031. 4200011184500 SUPPLY AND INSTALLATION OF THE REQUIRED XLEF/PVC POWER CABLES, 3 PHASE, 380V, 185 MMZ CABLE WIRES RELATED TO FREEZER 68 BODY NMG10009-P034. 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG10009-P034. 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG10009-P034. 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG10009-P036. 420001118400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG10009-P036. 42000118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG10009-P036. 42000118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG10009-P038. 42000118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG1009-P039. 42000118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQLIPMENT ROOM. 250 AMP RELATED TO FREEZER 68 BODY NMG1009-P039. 42000118500 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 68 BODY NMG1009-P039. 42000118500 SUPPLY AND INSTALLATION OF THE A/C DIFFUSER'S COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 68 BODY NMG1009-P031. 42000118500 SUPPLY AND INSTALLATION OF THE A/C DIFFUSER'S COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HAAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER 68 BODY NMG10009-P031. 42000118500 SUPPLY AND INSTALLATION OF THE A/C DIFFUSER'S COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HAAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER 68 BODY NMG10009-P031. 42000118500 SUPPLY AND INSTALLATION OF THE A/C DIFFUSER'S COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AN				-
NMG1009-P032 4200011184700 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO FREEZER 06 BODY NMG10009-P034 4200011184700 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO FREEZER 06 BODY NMG10009-P034 420001118490 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 06 BODY NMG10009-P034 420001118490 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 06 BODY NMG10009-P037 420001118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 06 BODY NMG10009-P037 420001118500 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW), RELATED TO FREEZER 06 BODY NMG10009-P037 4200011185100 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW), RELATED TO FREEZER 06 BODY NMG10009-P039 420001118500 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS, RELATED TO FREEZER 06 BODY NMG10009-P040 420001118500 SUPPLY AND INSTALLATION OF THE ACCOUNTED WITH CABLES. AS PER HOSPITAL STANDARDS, RELATED TO FREEZER 06 BODY NMG10009-P040 420001118500 SUPPLY INSTALLATION OF THE ACCOUNTED WITH CABLES. AS PER HOSPITAL STANDARDS, RELATED TO FREEZER 06 BODY NMG10009-P040 420001118500 SUPPLY INSTALLATION OF THE ACCOUNTED WITH FROM THE ACCOUNTED WITH THE ACCOU				
NMG1009-P033 4200011184700 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO FREEZER 06 BODY NMG10009-P035 4200011184900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 06 BODY NMG10009-P036 4200011184900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 06 BODY NMG10009-P037 420001118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 06 BODY NMG10009-P037 420001118500 SUPPLY AND INSTALLATION OF AGRIE TRAY INSTALLATION OF AGRIE TRAY INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P039 420001118300 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P030 420001118300 SUPPLY AND INSTALLATION OF THE EIPHONE OUTLET COMPLETE WITH PLANUM BOX, FLEXRIBLE, COLAR AND AS PER HAVG DRAWING AND LOCATION. [WITHOUT HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P040 420001118500 SUPPLY INSTALLATION OF THE ACT DIFFERENCE OMPLETE WITH PLENUM BOX, FLEXRIBLE, COLAR AND AS PER HAVG DRAWING AND LOCATION. [WITH HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P041 420001118500 SUPPLY NATURALITATION OF THE ACT DIFFEREZER 06 BODY NMG10009-P042 420001118500 SUPPLY NATURALITATION OF THE ACT DIFFEREZER 06 BODY NMG10009-P044 420001118500 SUPPLY NATURALITATION OF THE ACT DIFFEREZER 06 BODY NMG10009-P044 420001118500 SUPPLY NATURALITATION OF THE ACT DIFFEREZER 06 BODY NMG10009-P045 420001118500 SUPPLY NATURALITATION OF THE ACT DIFFEREZER 06 BODY NMG10009-P046 420001118500 SUPPLY AND INSTALLATION OF ACC DIFFEREZER 06 BODY NMG10009-P047 420001118500 SUPPLY AND INSTALLATION OF ACC DIFFEREZER 06 BODY NMG10009-P047 420001118500 SUPPLY AND INSTALLATION OF ACC DIFFEREZER 06 BODY NMG10009-P047 420001118500 SUPPLY AND INSTALLATION OF ACC DIFFEREZER 06 BODY NMG10009-P047 420001118500 SUPPLY AND INSTALLATION OF ACC	NMG10009-P031	4200011184500	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P034 4200011184900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 06 BODY NMG10009-P035 420001118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 06 BODY NMG10009-P036 420001118500 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 06 BODY NMG10009-P037 420001118500 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 06 BODY NMG10009-P039 420001118500 SUPPLY AND INSTALLATION OF FOWER OUTLIET COMPLETE WITH HAPS. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P039 420001118500 SUPPLY AND INSTALLATION OF THE PHONE OUTLIET COMPLETE WITH HAPS. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P034 420001118500 SUPPLY AND INSTALLATION OF THE FALSE COMPLETE WITH PLEAVIN BOX, FLEXIBLE, CLORA AND AS PER HAVO EARNING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER MG10009 NMG10009-P04 420001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLEAVIN BOX, FLEXIBLE, CLORA AND AS PER HAVO EARNING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER MG10009 NMG10009-P04 420001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLEAVIN BOX, FLEXIBLE, CLORA AND AS PER HAVO EARNING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER MG10009 NMG10009-P04 420001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLEAVIN BOX, FLEXIBLE, CLORA AND AS PER HAVE DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER OF BODY NMG10009-P04 420001118500 SUPPLY AND INSTALLATION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAW PIPES, NMG10009-P04 A20001118500 SUPPLY AND INSTALLATION OF A/C DIFFUSER (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER OF BODY NMG10009-P04 A20001118500 SUPPLY AND INSTALLATION OF A/C DIFFUSER (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER OF BODY NMG10009-P04	NMG10009-P032			NMG10009
NMG1009-P035 420001118500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 06 BODY NMG10009-P036 420001118500 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP). COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW NMG10009-P037 4200011185100 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FAISE CEILING (KW), RELATED TO FREEZER 06 BODY NMG10009-P038 420001118500 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P039 120001118500 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH LAS ALTED TO FREEZER 06 BODY NMG10009-P040 420001118500 SUPPLY INSTALLATION OF THE PA/C DIFFUSERS COMPLETE WITH LAS ALTED TO FREEZER OF BODY NMG10009-P041 420001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER OF MMG10009-P049 120001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH INTO THE PA/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER OF MMG10009-P049 120001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER OF MMG10009-P049 120001118500 FIREWATER LINE RELATED TO FREEZER OF BODY NMG10009-P049 120001118500 FIREWATER LINE RELATED TO FREEZER OF BODY NMG10009-P049 120001118500 FIREWATER LINE RELATED TO FREEZER OF BODY NMG10009-P049 120001118500 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN), (WITHOUT HEPA FILTER) RELATED TO FREEZER OF BODY NMG10009-P049 120001118500 SUPPLY AND INSTALLATION OF A/C GRILLS, (WITH HEPA FILTER) RELATED TO FREEZER OF BODY NMG10009-P049 120001118500 SUPPLY AND INSTALLATION OF A/C GRILLS, (WITH HEPA FILTER) RELATED TO FREEZER OF BODY NMG10009-P050 120001118500 SUPPLY AND INST	NMG10009-P033	4200011184700	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P034 420001118500 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW (NMG10009-P037) 4200011185100 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 06 BODY NMG10009-P038 4200011185300 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P040 4200011185300 SUPPLY AND INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P041 4200011185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 (NMG10009-P041 4200011185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 (NMG10009-P041 420001118500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 (NMG10009-P041 420001118500 SUPPLY INSTALLATION OF FREEZER 06 BODY MMG10009-P041 420001118500 SUPPLY SUPP	NMG10009-P034	4200011184800	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P037 420011185100 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW), RELATED TO FREEZER OB BODY NMG10009-P039 4200011185300 SUPPLY AND INSTALLATION OF THE AVENT COMPLETE WITH PLASHES, AS PER HOSPITAL STANDARDS, RELATED TO FREEZER OB BODY NMG10009-P040 4200011185400 SUPPLY AND INSTALLATION OF THE AVENT COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HAVE DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P041 420001118500 SUPPLY INSTALLATION OF THE AVENT COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HAVE DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P042 420001118500 SUPPLY INSTALLATION OF THE AVENT COMPLETE WITH PLENUM ROW, FLEXIBLE, COLAR AND AS PER HAVE DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P043 420001118500 SUPPLY INSTALLATION OF THE AVENT COMPLETE WITH PLENUM ROW, FLEXIBLE, COLAR AND AS PER HAVE DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P044 420001118500 SUPPLY AND INSTALLATION OF THE AVENT COMPLETE WITH PLENUM ROW, FLEXIBLE, COLAR AND AS PER HAVE DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER NMG10009 NMG10009-P044 420001118500 SUPPLY AND INSTALLATION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAWN PIPES, NMG10009 NMG10009-P047 420001118500 SPRINGKLER HEADS RELATED TO FREEZER OB BODY NMG10009 NMG10009-P047 420001118500 SPRINGKLER HEADS RELATED TO FREEZER OB BODY NMG10009-P047 420001118500 SUPPLY AND INSTALLATION OF AVENT COMPLETE WITH PLEASE PLEATED TO FREEZER OB BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF AVENT COMPLETE WITH PLEASE PLEATED TO FREEZER OB BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF AVENT COMPLETE WITH PLEASE PLEATED TO FREEZER OB BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF AVENT COMPLETE WITH PLEASE PLEATED TO FREEZER OB BODY NMG10009-P049 42000118500 SUPPLY AND INSTALL	NMG10009-P035	4200011184900	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P038 420001185200 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P040 4200011185300 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RIJS. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009-P041 4200011185300 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 NMG10009 NMG10009-P042 4200011185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 NMG10009 NMG10009-P043 4200011185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 NMG10009 NMG10009-P044 420001118500 FIRE WATER LINE RELATED TO FREEZER 06 BODY NMG10009-P045 420001118500 SUPPLY SIDE STANDARD FIRE WATER LINE RELATED TO FREEZER 06 BODY NMG10009-P046 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P047 40001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P048 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P049 420001118500 SUPPLY AND INSTALLATION OF A/C DUCTS.	NMG10009-P036	4200011185000	SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW	/ NMG10009
NMG10009-P039 420001185300 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH R145. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P040 420001185400 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 MNG10009- NMG10009-P042 420001185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 MNG10009- NMG10009-P042 420001185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 06 MNG10009- NMG10009-P044 420001185500 RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPES, NMG10009- NMG10009-P045 A20001185900 DRAIN PIPES RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P046 A20001185900 DRAIN PIPES RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P046 A20001185000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P047 A20001185000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P048 A20001186300 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P048 A20001186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P044 A20001186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 A20001186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 A20001186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 A20001186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 0	NMG10009-P037	4200011185100	SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P041 420011185400 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 MMG10009 NMG10009-P042 4200011185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 6 MMG10009 NMG10009-P043 420011185500 FIRE WATER LINE RELATED TO FREEZER 6 BODY NMG10009-P044 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-P045 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-P045 A2001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 6 BODY NMG10009-P046 A2001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P048 A2001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P049 A2001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P054 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P054 A200	NMG10009-P038	4200011185200	SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P041 420011185400 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 MMG10009 NMG10009-P042 4200011185500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 6 MMG10009 NMG10009-P043 420011185500 FIRE WATER LINE RELATED TO FREEZER 6 BODY NMG10009-P044 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-P045 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-P045 A2001118500 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 6 BODY NMG10009-P046 A2001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P048 A2001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P049 A2001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P051 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P054 A20001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY) RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-P054 A200	NMG10009-P039	4200011185300	SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RJ45. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 06 BODY	NMG10009
NMG1009-P042 42001118500 FIRE WATER LINE RELATED TO FREEZER 66 BODY NMG10009-NMG10009-P043 42001118500 FIRE WATER LINE RELATED TO FREEZER 6 BODY NMG10009-NMG10009-P044 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-NMG10009-P045 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-NMG10009-P045 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P046 42001118500 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P047 420011186100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P048 420011186200 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118630 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-N	NMG10009-P040	4200011185400	SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZE	RNMG10009
NMG1009-P042 42001118500 FIRE WATER LINE RELATED TO FREEZER 66 BODY NMG10009-NMG10009-P043 42001118500 FIRE WATER LINE RELATED TO FREEZER 6 BODY NMG10009-NMG10009-P044 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-NMG10009-P045 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-NMG10009-P045 42001118500 SPRINGKLER HEADS RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P046 42001118500 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P047 420011186100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P048 420011186200 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118630 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-P050 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 6 BODY NMG10009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-NMG1009-N				
NMG1009-P043 420011185700 FIRE WATER LINE RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P044 4200011185900 DRAIN PIPES RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P045 4200011185000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P046 420001118600 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P047 4200011186100 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P049 420001118600 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P049 4200011186300 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009- NMG1009-P049 4200011186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P050 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P051 4200011186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P052 4200011186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P053 4200011186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 4200011186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 4200011186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009- NMG10009-W3 42000118600 SUPPLY A		+	, , ,	
NMG1009-P044 420011185800 SPRINGKLER HEADS RELATED TO FREEZER 06 BODY NMG1009-P045 4200011185900 DRAIN PIPES RELATED TO FREEZER 06 BODY NMG10009-P046 4200011186900 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P048 420001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P049 420001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 420001118600 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P053 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P055 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P056 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P056 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P056 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P056 42000118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P056 42000118600		-		4
NMG1009-P045 420011185900 DRAIN PIPES RELATED TO FREEZER 06 BODY NMG1009-P046 420011185900 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG1009-P047 420001186300 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P048 420001186300 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P049 420001186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P049 420001186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 A20001186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 A20001186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 A20001186600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P054 A20001186600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P054 A20001186600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P054 A20001186600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P055 A20001186600 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FILODOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P056 A20001186900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FILODOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P056 A20001186900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FILODOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P056 A20001186900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FILODOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED				
NMG1009-P044 42001118600 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 06 BODY NMG10009-P048 420011186100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P048 420011186200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P049 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P052 420001118600 SUPPLY OF PORTABLE COZ FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P053 420001118600 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P054 420001118600 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WERE AND				
NMG1009-P047 420011186100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P048 420011186200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P049 420011186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 4200011186500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZER NMG10009-P051 4200011186600 SUPPLY OF PORTABLE COZ FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P052 4200011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P054 420001118690 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P054 420001118690 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-P056 420001118690 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-P056 420001118690 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-P056 420001118500 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-P056 A20001118500 SUPPLY AND INSTALLATED TO FREEZER 06 BODY NMG10009-P056 A20001118500 SUPPLY AND INSTALLATED TO FREEZER 06 BODY NMG10009-P056 A200011653300 SIT YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-P056 SUPPLY AND INSTALLATED TO FREEZER 06 BODY NMG10009-P056 SUPPLY AND INSTALLATED TO FREEZER 06 BODY NMG10009-P056 SUPPLY AND INSTA		-		
NMG10009-P048 4200011186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 4200011186500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 4200011186500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZER NMG10009 NMG10009-P052 4200011186500 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P053 4200011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P054 4200011186800 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P055 4200011186900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P056 4200011186900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-P056 4200011187000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 4200011653300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 4200011653300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 4200011653300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 4200011653300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 420001165300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 420001165300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 420001065300 STRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 5TRY YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10			, , , , , , , , , , , , , , , , , , ,	
NMG10009-P049 4200011186300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P050 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 4200011186500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZER NMG10009 NMG10009-P052 4200011186600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P053 4200011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P054 4200011186800 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P055 4200011186900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 4200011853300 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 06 BODY NMG10009-W4 4200011653300 3rd YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653300 Sth YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009				
NMG10009-P050 4200011186400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 06 BODY NMG10009-P051 4200011186500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZEF, NMG10009 NMG10009-P052 420001118600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P053 4200011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER, NMG10009 NMG10009-P054 4200011186800 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FO NMG10009 NMG10009-P055 4200011187000 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 4200011653300 3rd YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W4 420011653300 3rd YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W5 420011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W6 420610653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY		-		
NMG10009-P051 4200011186500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZER NMG10009-NMG10009-P052 420001118600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P053 4200011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-NMG10009-P054 4200011186800 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FO NMG10009-NMG10009-P056 4200011187000 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-W15 4200011653300 3rd YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W15 4200011653400 4th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-NMG10009-W5 420011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-NMG				
NMG10009-P052 420011186600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 06 BODY NMG10009-P053 420011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P054 420011186800 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009 NMG10009-P055 420011186900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 420011187000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 06 BODY NMG10009-W4 420011653300 3rd YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420011653400 4th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420011653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420610653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420610653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420610653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420610653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420610653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W5 420610653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY				
NMG10009-P053 420011186700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-NMG10009-P054 420011186800 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER NMG10009-NMG10009-P054 420011186900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 420011187000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 420011653300 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653400 4th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 STA YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-				
NMG10009-P054 420011186800 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR MMG10009-P056 A20011186900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 420011187000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 06 BODY NMG10009-W3 420011653300 3rd YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653400 4th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 420011653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009-W3 4226180600100 FREEZER 06 BODY NMG10009-W3 4226180600100 FREEZER 06 BODY NMG10009-W3 4226180600100 FREEZER 06 BODY				
NMG10009-P055 4200011186900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 06 BODY NMG10009-NMG10009-P056 MMG10009-P056 MMG10009-P056 <td></td> <td></td> <td></td> <td></td>				
NMG10009-P056 4200011187000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 06 BODY NMG10009 NMG10009-W3 4200011653300 3rd YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009 NMG10009-W4 4200011653400 4th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009 NMG10009-W5 4200011653500 5th YEAR WARRANTY COST RELATED TO FREEZER 06 BODY NMG10009 NMG10009 4226180600100 FREEZER 06 BODY NMG10009				_
NMG10009-W3 4200011653300 3rd YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 4200011653400 4th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3 NMG10009-W3 4200011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009-W3				
NMG10009-W4 420011653400 4th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009 NMG10009-W5 420011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009 NMG10009 4226180600100 FREEZER 06 BODY NMG10009				+
NMG10009-W5 420011653500 5th YEAR WARRANTY COST RELATED TOFREEZER 06 BODY NMG10009 NMG10009 4226180600100 FREEZER 06 BODY NMG10009		-		_
NMG10009 4226180600100 FREEZER 06 BODY NMG10009				
	NMG10009-W5			+
NMG10010-001 4226180602200 CENTRAL STATION CONTROL RELATED TO FREEZER 09 BODY NMG10010	NMG10009			
	NMG10010-001	4226180602200	CENTRAL STATION CONTROL RELATED TO FREEZER 09 BODY	NMG10010

NMG10010-002 4226180602300 RECORDER RELATED TO FREEZER 09 BODY NMG10010-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO FREEZER 09 BODY NMG10010-D2 400000054400 REMOTE AREA DELIVERY COST RELATED TO FREEZER 09 BODY NMG10010-P001 4200011187100 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BU NMG10010-P002 4200011187200 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO FREEZER 09 BODY	NMG10010 NMG10010
NMG10010-D2 400000054400 REMOTE AREA DELIVERY COST RELATED TO FREEZER 09 BODY NMG10010-P001 4200011187100 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BU	
NMG10010-P001 4200011187100 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BU	NMG10010
	NMG10010
NMG10010-P003 4200011187300 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS	
NMG10010-P004 4200011187400 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P005 4200011187500 SUPPLY AND INSTALLATION OF CABLE THAT RELATED TO TREEZER 09 BODY	NMG10010
NMG10010-P006 4200011187600 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA. RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P000 4200011187700 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P008 4200011187800 REMOVAL, DEMOLITION OF EXISTING WILDIGAL EQUIPMENT. WORKS INCLUDE, WOOD CHATTING AND THANSFORT TO THOSE THE SWARE HOUSE. RELATED TO TREEZER 09 BODY	NMG10010
NMG10010-P009 4200011187900 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO FREEZER 09 BODY NMG10010-P010 4200011188000 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO FR	NMG10010
, , ,	
NMG10010-P012 4200011188200 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER 0	
NMG10010-P013 4200011188300 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. M	
NMG10010-P014 4200011188400 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLE	
NMG10010-P015 4200011188500 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVE	
NMG10010-P016 4200011188600 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P017 4200011188700 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P018 4200011188800 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P019 4200011188900 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P020 4200011189000 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P021 4200011189100 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P022 4200011189200 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P023 4200011189300 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BI	REA NMG10010
NMG10010-P024 4200011189400 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P025 4200011189500 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P026 4200011189600 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P027 4200011189700 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P028 4200011189800 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P029 4200011189900 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P030 4200011190000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P031 4200011190100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P032 4200011190200 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P033 4200011190300 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P034 4200011190400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P035 4200011190500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P036 4200011190600 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLC	
NMG10010-P037 4200011190700 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P038 4200011190800 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P039 4200011190900 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RI45. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P040 4200011191000 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREE	_
NMG10010-P041 4200011191100 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER NMG10010-P042 4200011191200 RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIFES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT.	
NMG10010-P042 4200011191200 RELOCATION/ DIVERSION OF ANY EXISTING OTHER FIRE WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SOCH AS FIRE WATER LINE, SPRINGREER HEADS, DRAIN FIRE NAME OF ANY EXISTING OTHER FIRE WATER LINE, SPRINGREER HEADS, DRAIN FIRE WATER LINE, SPRI	
	NMG10010
NMG10010-P044 4200011191400 SPRINGKLER HEADS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P045 4200011191500 DRAIN PIPES RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P046 4200011191600 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P047 4200011191700 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P048 4200011191800 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P049 4200011191900 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P050 4200011192000 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P051 4200011192100 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREE	
NMG10010-P052 4200011192200 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 09 BODY	NMG10010

NMG10010-P053 4200011192300 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FRE	EZER NMG10010
NMG10010-P054 4200011192400 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DAT	
NMG10010-P055 4200011192500 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-P056 4200011192600 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 09 BODY	NMG10010
NMG10010-W3 4200011653700 3rd YEAR WARRANTY COST RELATED TOFREEZER 09 BODY	NMG10010
NMG10010-W4 4200011653800 4th YEAR WARRANTY COST RELATED TOFREEZER 09 BODY	NMG10010
NMG10010-W5 4200011653900 5th YEAR WARRANTY COST RELATED TOFREEZER 09 BODY	NMG10010
NMG10010 4226180600200 FREEZER 09 BODY	NMG10010
NMG10011-001 4226180602400 CENTRAL STATION CONTROL RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-002 4226180602500 RECORDER RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-D1 400000054300 MAIN AREA DELIVERY COST RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-D2 400000054400 REMOTE AREA DELIVERY COST RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P001 4200011192700 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE	
NMG10011-P002 4200011192800 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P003 4200011192900 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR A	
NMG10011-P004 4200011193000 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P005 4200011193100 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P006 4200011193200 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P007 4200011193300 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P008 4200011193400 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P009 4200011193500 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO FREEZER 12 BOD	
NMG10011-P010 4200011193600 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO I	
NMG10011-P011 4200011193700 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER	
NMG10011-P012 4200011193800 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER	
NMG10011-P013 4200011193900 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING.	
NMG10011-P014 4200011194000 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERA	
NMG10011-P015 4200011194100 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVERN WELDED SEAMS EXTENDED TO THE WALL WE WALL WE WALL WE WANT WITH WELD SEAMS EXTENDED TO THE WALL WE WALL WE WANT WAS AND WE WA	
NMG10011-P016 4200011194200 SUPPLY AND INSTALLARIO GOOD SYSTEM RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P017 4200011194300 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P018 4200011194400 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO TREEZER 12 BODY	NMG10011
NMG10011-P019 4200011194500 SUPPLY AND INSTALL STRINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P020 4200011194500 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO TREEZER 12 BODY	NMG10011
NMG10011-P020 420011194000 30FFET AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS FER HOSPITAL REQUIREMENTS RELATED TO FREEZER 12 BODY NMG10011-P021 4200011194700 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P022 4200011194700 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO TREEZER 12 BODY	NMG10011
NMG10011-P022 4200011194900 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO PREEZER 12 BODT NMG10011-P023 4200011194900 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT	
NMG10011-P024 4200011195000 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P025 4200011195100 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P026 4200011195200 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO FREEZER 12 BODY	NMG10011 NMG10011
NMG10011-P027 4200011195300 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES, 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO FREEZER 12 BODY	NMG10011 NMG10011
	NMG10011 NMG10011
NMG10011-P029 4200011195500 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P030 4200011195600 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P031 4200011195700 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P032 4200011195800 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P033 4200011195900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P034 4200011196000 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P035 4200011196100 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P036 4200011196200 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW B	
NMG10011-P037 4200011196300 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P038 4200011196400 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P039 4200011196500 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RJ45. AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P040 4200011196600 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FR	
NMG10011-P041 4200011196700 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZE	
NMG10011-P042 4200011196800 RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN F	
NMG10011-P043 4200011196900 FIRE WATER LINE RELATED TO FREEZER 12 BODY	NMG10011

NMG10011-P044 4200011197000 SPRINGKLER HEADS RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P045 4200011197100 DRAIN PIPES RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P046 4200011197200 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P047 4200011197300 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P048 4200011197400 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P049 4200011197500 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P050 4200011197600 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P051 4200011197700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FRE	
NMG10011-P052 4200011197800 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P053 4200011197900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREE.	
NMG10011-P054 4200011198000 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA	
NMG10011-P055 4200011198100 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-P056 4200011198200 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO FREEZER 12 BODY	NMG10011
NMG10011-W3 4200011654100 3rd YEAR WARRANTY COST RELATED TOFREEZER 12 BODY	NMG10011
NMG10011-W4 4200011594200 4th YEAR WARRANTY COST RELATED TOFREEZER 12 BODY	NMG10011
NMG10011-W5 4200011654300 Sth YEAR WARRANTY COST RELATED TOFREEZER 12 BODY	NMG10011
NMG10011 4226180600300 FREEZER 12 BODY	NMG10011
NMG10012-001 4226180602600 CENTRAL STATION CONTROL RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-002 4226180602700 RECORDER RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-012 400000054300 MAIN AREA DELIVERY COST RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-D2 400000054400 REMOTE AREA DELIVERY COST RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-D2 400000034400 REMOTE AREA DELIVERY COST RELATED TO TREEZER 21 BODT NMG10012-P001 4200011198300 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BU	
NMG10012-F001 4200011198000 FREPARING THE SHE TO START WORK BY REMOVING, DISMARTEINO, CRACKING AND MOVING ALE THAT IS NEEDED TO DISMARTEE WALLS, FLOORS AND CETEINGS FROM INSIDE THE BO NMG10012-F002 4200011198400 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-F002 4200011198500 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS I	
NMG10012-P003 4200011198000 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-F004 420011198700 SUPPLY AND INSTALLATION OF CABLE TRAIN RELATED TO TREEZER 21 BODY NMG10012-P005 4200011198700 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P005 4200011198700 SUPPER AND INSTALLATION OF PARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO PREEZER 21 BODY NMG10012-P006 4200011198800 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA RELATED TO PREEZER 21 BODY	NMG10012
NMG10012-P000 4200011198900 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P007 4200011199000 DISMANTEING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLODE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO PREEZER 21 BODY NMG10012-P008 4200011199000 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO PREEZER 21 BODY	NMG10012
NMG10012-P008 4200011199000 REMOVAL, DEMOCRITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO FREEZER 21 BODY NMG10012-P009 4200011199100 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P009 4200011199100 REMOVAL AND RE	
NMG10012-P010 4200011199200 CONSTROCTION OF MASONAY WALLS WITH FIRE RATED AND MOISTORE RESISTANT, FIGER OR ROCKWOOD SOUND-PROOF INSOLATION, AS REQUIRED IN THE DRAWING. RELATED TO FREEZER 21 NMG10012-P011 4200011199300 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER 21	
NMG10012-P011 4200011199300 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER 21 NMG10012-P012 4200011199400 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO FREEZER 21	
NMG10012-P012 4200011199500 SUPPLY AND INSTALLATION OF MICHORIZED SCIDING DOOR, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. RELATED TO FREEZER 2. NMG10012-P013 4200011199500 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. M.	
	_
NMG10012-P014 4200011199600 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABL NMG10012-P015 4200011199700 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVE	
	_
NMG10012-P016 4200011199800 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P017 4200011199900 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P018 4200011200000 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P019 4200011200100 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P020 4200011200200 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P021 4200011200300 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P022 4200011200400 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO FREEZER 21 BODY	NMG10012
INMG10012-P023 4200011200500 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BE	
NMG10012-P024 4200011200600 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P025 4200011200700 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P026 4200011200800 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P027 4200011200900 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P028 4200011201000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P029 4200011201100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P030 4200011201200 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P031 4200011201300 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P032 4200011201400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM 200 AMP RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P033 4200011201500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-F033 4-200011201500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 200 AMP RELATED TO FREEZER 21 BODY	NMG10012

NMG10012-P035	4200011201700 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P036	4200011201800 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW F	
NMG10012-P037	4200011201900 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P038		NMG10012
NMG10012-P039		NMG10012
NMG10012-P040	4200011202200 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO FREEZER	
NMG10012-P041	4200011202300 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO FREEZER 21 B	
NMG10012-P042		NMG10012
NMG10012-P043	4200011202500 FIRE WATER LINE RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P044		NMG10012
NMG10012-P045		NMG10012
NMG10012-P046		NMG10012
NMG10012-P047		NMG10012
NMG10012-P048	, , , , , , , , , , , , , , , , , , , ,	NMG10012
NMG10012-P049		NMG10012
NMG10012-P050		NMG10012
NMG10012-P051	4200011203300 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO FREEZER	
NMG10012-P052	4200011203400 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO FREEZER 21 BODY	NMG10012
NMG10012-P053	4200011203500 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO FREEZER 2	
NMG10012-P054	4200011203600 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR	
NMG10012-P055		NMG10012
NMG10012-P056		NMG10012
NMG10012-F030		NMG10012
NMG10012-W3		NMG10012
NMG10012-W4		NMG10012
NMG10012-W3		NMG10012
NMG10012 NMG10016-001		NMG10012 NMG10016
NMG10016-001		NMG10016
NMG10016-D1		NMG10016
NMG10016-W3		NMG10016
NMG10016-W4		NMG10016
NMG10016-W4		NMG10016
NMG10016-W3		NMG10016
NMG10017-D1		NMG10017
NMG10017-D1 NMG10017-D2		NMG10017 NMG10017
NMG10017-D2		NMG10017 NMG10017
NMG10017-W3		NMG10017 NMG10017
NMG10017-W4		NMG10017 NMG10017
NMG10017-W3		NMG10017 NMG10017
NMG10017		NMG10017 NMG10019
NMG10019-001		NMG10019 NMG10019
NMG10019-002		NMG10019 NMG10019
NMG10019-D1		NMG10019 NMG10019
NMG10019-D2 NMG10019-P001		
NMG10019-P001	4200011203900 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILDI 4200011204000 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO REFRIGERATOR 03 BODY	
		NMG10019
NMG10019-P003	4200011204100 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER	
NMG10019-P004		NMG10019
NMG10019-P005	4200011204300 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO REFRIGERATOR 03 BODY	
NMG10019-P006		NMG10019
NMG10019-P007	4200011204500 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P008		NMG10019
NMG10019-P009	4200011204700 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL) WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO REFRIGERATOR 03 BOX	
NMG10019-P010	4200011204800 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO REFRIG	
NMG10019-P011	4200011204900 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING, MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATOR	
MIVIG 10019-P012	4200011205000 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATOR	NIVIG10019

NMG10019-P013 420001120	100 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME SIZE AS PER DRAWING. MAT	FINMG10019
	SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY	
	SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER	
	3400 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P017 420001120		NMG10019
NMG10019-P018 420001120		NMG10019
NMG10019-P019 420001120		NMG10019
NMG10019-P020 420001120		NMG10019
NMG10019-P021 420001120		NMG10019
	SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P023 420001120	·	
NMG10019-P024 420001120		NMG10019
	3300 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P026 420001120	, , ,	NMG10019
NMG10019-P027 420001120		NMG10019
	5600 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P029 420001120		NMG10019
NMG10019-P030 420001120		NMG10019
	1900 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P032 420001120		NMG10019
NMG10019-P032 420001120		NMG10019
	200 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P035 420001120		NMG10019
NMG10019-P036 420001120		
	SOO SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO REFRIGERATOR 03 BODY	NMG10019
	7600 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 03 BODY 7700 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RI45. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 03 BODY	NMG10019 NMG10019
	1800 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO REFRIG	
NMG10019-P041 420001120 NMG10019-P042 420001120		
		_
	13100 FIRE WATER LINE RELATED TO REFRIGERATOR 03 BODY	NMG10019
		NMG10019 NMG10019
	DRAIN PIPES RELATED TO REFRIGERATOR 03 BODY	
	SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P047 420001120		NMG10019
NMG10019-P048 420001120		NMG10019
	STOOL SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P050 420001120	, , ,	NMG10019
NMG10019-P051 420001120		
	SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P053 420001120	,	
NMG10019-P054 420001120		
	3300 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 03 BODY	NMG10019
NMG10019-P056 420001120		NMG10019
NMG10019-W3 420001165		NMG10019
NMG10019-W4 420001165		NMG10019
NMG10019-W5 420001165		NMG10019
NMG10019 422618040		NMG10019
	1700 CENTRAL STATION CONTROL RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-002 422618040		NMG10020
NMG10020-D1 40000000		NMG10020
	1400 REMOTE AREA DELIVERY COST RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P001 420001120		
NMG10020-P002 420001120		NMG10020
NMG10020-P003 420001120	1700 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PEI	R NMG10020

	T.,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,
NMG10020-P004 4200011209800 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P005 4200011209900 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO REFRIGERATOR 06 BOD	
NMG10020-P006 4200011210000 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P007 4200011210100 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P008 4200011210200 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P009 4200011210300 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO REFRIGERATOR 06 BC	
NMG10020-P010 4200011210400 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO REFRI	
NMG10020-P011 4200011210500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATO	
NMG10020-P012 4200011210600 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATO	
NMG10020-P013 4200011210700 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME SIZE AS PER DRAWING. MATI	
NMG10020-P014 4200011210800 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY (
NMG10020-P015 4200011210900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER I	
NMG10020-P016 4200011211000 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P017 4200011211100 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P018 4200011211200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P019 4200011211300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P020 4200011211400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P021 4200011211500 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P022 4200011211600 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P023 4200011211700 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREA	
NMG10020-P024 4200011211800 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P025 4200011211900 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P026 4200011212000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P027 4200011212100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P028 4200011212200 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P029 4200011212300 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P030 4200011212400 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P031 4200011212500 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P032 4200011212600 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P033 4200011212700 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P034 4200011212800 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P035 4200011212900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P036 4200011213000 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW	FNMG10020
NMG10020-P037 4200011213100 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P038 4200011213200 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P039 4200011213300 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH R.45. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P040 4200011213400 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO REFRIGI	E NMG10020
NMG10020-P041 4200011213500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO REFRIGERATE	ī(NMG10020
NMG10020-P042 4200011213600 RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPES,	, NMG10020
NMG10020-P043 4200011213700 FIRE WATER LINE RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P044 4200011213800 SPRINGKLER HEADS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P045 4200011213900 DRAIN PIPES RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P046 4200011214000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P047 4200011214100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P048 4200011214200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P049 4200011214300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P050 4200011214400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P051 4200011214500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO REFRIG	E NMG10020
NMG10020-P052 4200011214600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P053 4200011214700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO REFRIGE	R NMG10020
NMG10020-P054 4200011214800 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FO	RNMG10020
NMG10020-P055 4200011214900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-P056 4200011215000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO REFRIGERATOR 06 BODY	NMG10020
NMG10020-W3 4200011656100 3rd YEAR WARRANTY COST RELATED TOREFRIGERATOR 06 BODY	NMG10020
NMG10020-W4 4200011656200 4th YEAR WARRANTY COST RELATED TOREFRIGERATOR 06 BODY	NMG10020

NMG10020-W5	4200011656300 5th YEAR WARRANTY COST RELATED TOREFRIGERATOR 06 BODY	NMG10020
NMG10020	4226180400100 REFRIGERATOR 06 BODY	NMG10020
NMG10021-001	4226180400900 CENTRAL STATION CONTROL RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-002	4226180401000 RECORDER RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-D1	400000054300 MAIN AREA DELIVERY COST RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-D2	400000054400 REMOTE AREA DELIVERY COST RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P001	4200011215100 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILDI	
NMG10021-P002	4200011215200 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P003	4200011215300 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER	
NMG10021-P004	4200011215400 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO REFIGERATOR 09 BODY	NMG10021
NMG10021-P005	4200011215500 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO REFRIGERATOR 09 BODY	
NMG10021-P006	4200011215600 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA RELATED TO REFRIGERATION OP BODY	NMG10021
NMG10021-P007	4200011215700 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P008	4200011215800 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO REFRIGERATOR 99 BODY	NMG10021
NMG10021-P009	4200011215900 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL) / WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO REFRIGERATOR 09 BOI	
NMG10021-P009	4200011216000 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO REFRIG	
NMG10021-P010	4200011216100 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATOR	
NMG10021-P011	4200011216100 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATOR BY HOSPITAL RELATED TO REPRISE BY HOSPITAL RELATED TO REFRIGERATOR BY HOSPITAL RELATED TO REPRISE BY HOSPITAL RELATED BY HOSPITAL RELATED BY HOSPITAL RELATED B	
NMG10021-P012		
NMG10021-P013	4200011216300 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE	
	4200011216400 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY C	
NMG10021-P015	4200011216500 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F	
NMG10021-P016	4200011216600 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P017	4200011216700 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P018	4200011216800 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P019	4200011216900 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P020	4200011217000 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P021	4200011217100 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P022	4200011217200 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P023	4200011217300 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREA	
NMG10021-P024	4200011217400 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P025	4200011217500 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P026	4200011217600 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P027	4200011217700 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P028	4200011217800 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P029	4200011217900 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P030	4200011218000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P031	4200011218100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P032	4200011218200 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P033	4200011218300 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P034	4200011218400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P035	4200011218500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P036	4200011218600 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW I	NMG10021
NMG10021-P037	4200011218700 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P038	4200011218800 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P039	4200011218900 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH R.45. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P040	4200011219000 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO REFRIGE	
NMG10021-P041	4200011219100 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO REFRIGERAT	
NMG10021-P042		NMG10021
NMG10021-P043	4200011219300 FIRE WATER LINE RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P044	4200011219400 SPRINGKLER HEADS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P045	4200011219500 DRAIN PIPES RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P046	4200011219600 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P047	4200011219700 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P047	4200011219800 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P048	4200011219900 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO REFRIGERATOR 09 BODY	NMG10021
14141010021-F049	T200011213000 SOLLET AND INSTALLATION OF AYC GIVILD. (WITHOUT HERATELETIN) RELATED TO REININGERATION US DUT	LAINIO TOOST

NMG10021-P050	4200011220000 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P050	4200011220100 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO REFRIGE	
NMG10021-P052	4200011220200 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P053	4200011220300 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO REFRIGER	
NMG10021-P054	4200011220400 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR	
NMG10021-P055	4200011220500 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-P056	4200011220600 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO REFRIGERATOR 09 BODY	NMG10021
NMG10021-W3	4200011656500 3rd YEAR WARRANTY COST RELATED TOREFRIGERATOR 09 BODY	NMG10021
NMG10021-W4	4200011656600 4th YEAR WARRANTY COST RELATED TOREFRIGERATOR 09 BODY	NMG10021
NMG10021-W5	4200011656700 5th YEAR WARRANTY COST RELATED TOREFRIGERATOR 09 BODY	NMG10021
NMG10021-W3	4226180400200 REFRIGERATOR 09 BODY	NMG10021
NMG10022-001	4226180401100 CENTRAL STATION CONTROL RELATED TO REFRIGERATOR 12 BODY	NMG10021
NMG10022-001	4226180401200 RECORDER RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-002	400000054300 MAIN AREA DELIVERY COST RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-D1	400000054300 INMIN AREA DELIVERY COST RELATED TO REFRIGERATOR 12 BODY 400000054400 REMOTE AREA DELIVERY COST RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-D2	4200011220700 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILDI	
NMG10022-P001	4200011220800 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P003	4200011220900 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER	
NMG10022-P004	4200011221000 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P005	4200011221100 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO REFRIGERATOR 12 BODY	
NMG10022-P006	4200011221200 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P007	4200011221300 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P008	4200011221400 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P009	4200011221500 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO REFRIGERATOR 12 BOIL	
NMG10022-P010	4200011221600 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO REFRIG	
NMG10022-P011	4200011221700 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATOR	
NMG10022-P012	4200011221800 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATO	NMG10022
NMG10022-P013	4200011221900 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME SIZE AS PER DRAWING. MATE	NMG10022
NMG10022-P014	4200011222000 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY O	
NMG10022-P015	4200011222100 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F	NMG10022
NMG10022-P016	4200011222200 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P017	4200011222300 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P018	4200011222400 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P019	4200011222500 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P020	4200011222600 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P021	4200011222700 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P022	4200011222800 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P023	4200011222900 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREAI	NMG10022
NMG10022-P024	4200011223000 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P025	4200011223100 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P026	4200011223200 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P027	4200011223300 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P028	4200011223400 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P029	4200011223500 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P030	4200011223600 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P031	4200011223700 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P032	4200011223800 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM 200 AMP RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P033	4200011223900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P034	4200011224000 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 300 AMP RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P035	4200011224100 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM 400 AMP RELATED TO REFRIGERATOR 12 BODY	NMG10022
NMG10022-P035	4200011224200 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW F	
NMG10022-P030	4200011224300 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MIDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCOTT BREAKEN, SLOW BLOW F	NMG10022
NMG10022-P037		NMG10022
	4200011224400 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 12 BODY 4200011224500 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RJ45. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 12 BODY	
NMG10022-P039 NMG10022-P040	4200011224500 SUPPLY AND INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO REFRIGE	NMG10022
MINIG 10022-P040	14200011224000 30FFEL INSTALLATION OF THE AYOUTFUSEAS CONFEELE WITH FLENDIN BOA, FLEATBLE, COLAR AND AS PER HAVE DRAWING AND LOCATION. (WITHOUT HEPA FILLER) RELATED TO REPRIDE	IMINIGTOOSS

NMG10022-P041 4200	0011224700 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO REFRIGERAT	NMG10022
		NMG10022
	2011225700 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO REFRIGE	
		NMG10022
	1011225900 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO REFRIGER	
	1011225900 BE SUPPLYING AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE STSTEM WHICH INCLUDE FIFTING WORKS, TAFFED TO THE NEAREST BUILDING UTILITY STSTEM. RELATED TO REFRIGER 10011226000 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR	
		NMG10022
		NMG10022
	·	NMG10022
		NMG10022
		NMG10022
		NMG10022
		NMG10023
		NMG10023
NMG10023-D1 4000	000054300 MAIN AREA DELIVERY COST RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-D2 4000	0000054400 REMOTE AREA DELIVERY COST RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P001 4200	1011226300 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILDI	NMG10023
NMG10023-P002 4200	0011226400 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P003 4200	2011226500 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER	NMG10023
NMG10023-P004 4200	0011226600 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P005 4200	0011226700 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P006 4200	0011226800 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P007 4200	0011226900 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P008 4200	0011227000 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO REFRIGERATOR 21 BODY	NMG10023
	0011227100 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO REFRIGERATOR 21 BOL	
	0011227200 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO REFRIG	NMG10023
	0011227300 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATOR	
	0011227400 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING, MATERIAL AS APPROVED BY HOSPITAL. RELATED TO REFRIGERATO	
	0011227500 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME SIZE AS PER DRAWING. MATE	
	0011227600 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY O	
	3011227700 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER FO	
		NMG10023
		NMG10023
		NMG10023
		NMG10023
	·	NMG10023
		NMG10023
		NMG10023
	1011228500 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREAD	
		NMG10023
		NMG10023
	· · · · · · · · · · · · · · · · · · · ·	NMG10023
		NMG10023
NMG10023-P028 4200	0011229000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P029 4200	0011229100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P030 4200	0011229200 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P031 4200	0011229300 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO REFRIGERATOR 21 BODY	NMG10023

NMG10023-P032 420001122940	00 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM 200 AMP RELATED TO REFRIGERATOR 21 BODY	NMG10023
	30 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO REFRIGERATOR 21 BODY 30 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM, 250 AMP RELATED TO REFRIGERATOR 21 BODY	
	30 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO REFRIGERATOR 21 BODY 30 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO REFRIGERATOR 21 BODY	NMG10023 NMG10023
NMG10023-P036 420001122980		
	O SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PARKEL BOARD (MOP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GIT, CIRCOTT BREAKER, SLOW BLOW TO SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF CABLE TRATTINSIDE THE FALSE CEILING (AW). RELATED TO REFRIGERATOR 21 BODY	NMG10023
		NMG10023
	00 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RI45. AS PER HOSPITAL STANDARDS. RELATED TO REFRIGERATOR 21 BODY 00 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO REFRIGE	
	SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO REFRIGERATION.	
	, , ,	NMG10023
		
	00 FIRE WATER LINE RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SPRINGKLER HEADS RELATED TO REFRIGERATOR 21 BODY	NMG10023
	DO DRAIN PIPES RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P051 420001123130		
	00 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO REFRIGERATOR 21 BODY	NMG10023
	00 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO REFRIGER	NMG10023
NMG10023-P054 420001123160	DO BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOR	NMG10023
NMG10023-P055 420001123170	00 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-P056 420001123180	00 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO REFRIGERATOR 21 BODY	NMG10023
NMG10023-W3 420001165730	00 3rd YEAR WARRANTY COST RELATED TOREFRIGERATOR 21 BODY	NMG10023
NMG10023-W4 420001165740	00 4th YEAR WARRANTY COST RELATED TOREFRIGERATOR 21 BODY	NMG10023
NMG10023-W5 420001165750	00 Sth YEAR WARRANTY COST RELATED TOREFRIGERATOR 21 BODY	NMG10023
NMG10023 422618040040	00 REFRIGERATOR 21 BODY	NMG10023
NMG10025-001 422615130000	00 CIRCULAR SAW BLADE Ø 75 TO 76 MM RELATED TO SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
	00 SEGMENTAL SAW BLADE 50 TO 51 MM RELATED TO SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
NMG10025-003 422615130020		NMG10025
	DO DOUBLE SEGMENTAL SAW BLADE 63 TO 64 MM RELATED TO SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
	00 CIRCULAR SAW BLADE Ø 50 TO 51 MM RELATED TO SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
NMG10025-006 422615130050	· · · · · · · · · · · · · · · · · · ·	NMG10025
	10 DUST BAG RELATED TO SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
	100 MAIN AREA DELIVERY COST RELATED TO SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
NMG10025-D2 40000005440		NMG10025
	00 3rd YEAR WARRANTY COST RELATED TO SAW AUTOPS' ELECTRIC WITH EXTRACTOR	NMG10025
	00 4th YEAR WARRANTY COST RELATED TOSAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
NMG10025-W5 420001165790		NMG10025
	00 SAW AUTOPSY ELECTRIC WITH EXTRACTOR	NMG10025
	00 MAIN AREA DELIVERY COST RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-D2 40000005440		NMG10026
	DO PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILD	
	00 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P003 420001123210		
NMG10026-P004 420001123220		NMG10026
	00 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P006 420001123240		NMG10026
NMG10026-P007 420001123250	· · · · · · · · · · · · · · · · · · ·	NMG10026
	00 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P009 420001123270		NMG10026
NMG10026-P010 420001123280		
NMG10026-P011 420001123290	SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCALE AUTOPS	NMG10026

NMG10026-P012	4200011233000	SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCALE AL	TOPNMG10026
NMG10026-P013	+	SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. N	
NMG10026-P014		SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERAB	
NMG10026-P015		SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COV	
NMG10026-P016		SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P017	4200011233500	SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P018		SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P019	4200011233700	SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P020	4200011233800	SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P021	4200011233900	SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P022		·	NMG10026
NMG10026-P023	4200011234100	SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT B	
NMG10026-P024		INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P025	4200011234300	INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P026	4200011234400	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES: 3 PHASE. 380V, 35 MM2 CABLE WIRES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P027		SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 THASE, 380V, 50 MM2 CABLE WIRES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P028	4200011234600	SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P029	4200011234700	SUPPLY AND INSTALLATION OF THE REQUIRED XLP E/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P030		SUPPLY AND INSTALLATION OF THE REQUIRED XLP (7) YOU OWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P031	4200011234900	SUPPLY AND INSTALLATION OF THE REQUIRED XLP (7) YOU OWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P032	4200011234300	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P033		SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P034	4200011235100	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P035	4200011235200	SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P036		SUPPLY AND INSTALLATION OF MAINS DISCONNECT SWITCH IN THE EQUIPMENT ROOMS. 400 AMP RELATED TO SCALE ACTIONS!	
NMG10026-P037	4200011235400	SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION FANCE BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BE	NMG10026
NMG10026-P037	4200011235500		NMG10026
NMG10026-P039	+	SUPPLY AND INSTALLATION OF FOWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P039	4200011235700	SUPPLY AND INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO SCALE	
NMG10026-P040	4200011235800		
NMG10026-P041		RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PI	
NMG10026-P042	4200011236000	FIRE WATER LINE RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P044	4200011236100	SPRINGKLER HEADS RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P044		DRAIN PIPES RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P045	4200011236300	SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P046	4200011236400	'	
			NMG10026
NMG10026-P048		SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P049	4200011236700	SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P050	4200011236800	SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P051		CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO SCA	
NMG10026-P052	4200011237000	SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P053	4200011237100	,	
NMG10026-P054		BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA	
NMG10026-P055	4200011237300	SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO SCALE AUTOPSY	NMG10026
NMG10026-P056	4200011237400		NMG10026
NMG10026-W3		3rd YEAR WARRANTY COST RELATED TOSCALE AUTOPSY	NMG10026
NMG10026-W4	4200011658200	4th YEAR WARRANTY COST RELATED TOSCALE AUTOPSY	NMG10026
NMG10026-W5	4200011658300		NMG10026
NMG10026		SCALE AUTOPSY	NMG10026
NMG10027-D1			NMG10027
NMG10027-D2		REMOTE AREA DELIVERY COST RELATED TO SCRUB STATON SINGLE BAY	NMG10027
NMG10027-P001		PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BI	
NMG10027-P002	4200011237600	SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO SCRUB STATON SINGLE BAY	NMG10027
NMG10027-P003			
NMG10027-P004	4200011237800	SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO SCRUB STATON SINGLE BAY	NMG10027

NMG10027-P005 4200011238000 CLEARING AND CLEANING OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO SCRUB STATON SINGLE NMG1002 NMG10027-P007 4200011238000 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA. RELATED TO SCRUB STATON SINGLE BAY NMG1002 NMG10027-P008 4200011238100 DISMANTLINIG OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO SCRUB STATON SINGLE BAY NMG1002 NMG10027-P009 4200011238200 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO SCRUB STATON SINGLE BAY NMG1002 NMG10027-P010 4200011238300 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO SCRUB STATON SINGL NMG1002 NMG10027-P011 4200011238400 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO SCRUB MING1002 NMG10027-P011 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P013 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P013 4200011238900 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P015 4200011238900 SUPPLY AND INSTALLATION OF MASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY ON NMG1002 NMG10027-P015 4200011239900 SUPPLY AND INSTALL SINGLE WORD ON SERIES SEALED TO SCRUB STATON SINGLE BAY NMG1002 NMG10027-P0
NMG10027-P007 4200011238100 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P008 4200011238200 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P010 4200011238400 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO SCRUB STATON SINGL NMG10027-P011 4200011238400 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWNING. RELATED TO SCRUB MG1002 NMG10027-P011 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P013 4200011238600 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF SINGLE SWING DOOR, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P014 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG1002 NMG10027-P016 420001123900 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL SINGLE WORLD DOOR SRELATED TO SCRUB STATON SINGLE BAY NMG1002 NMG10027-P019 4200011239400 SUPPLY AND INSTALL SINGLE WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG1002 NMG10027-P020 4200011239400 SUPPLY AND INSTALL STAI
NMG10027-P008 4200011238200 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P010 4200011238300 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO SCRUB STATON SINGL NMG10027-P010 4200011238400 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO SCRUB STATON SINGLO NMG10027-P011 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF MOTORIZED SUIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG1002 NMG10027-P014 4200011238800 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE NMG1002 NMG10027-P015 4200011238800 SUPPLY AND INSTALLATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY ON NMG10027-P016 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG1002 NMG10027-P016 420001123900 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL STAILL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL STAILL STAILLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESS
NMG10027-P009 4200011238300 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO SCRUB STATON SINGL NMG10027-P010 4200011238400 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO SCRUB STATON NMG10027-P011 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P014 4200011238700 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P014 4200011238800 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P014 4200011238800 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P015 420001123900 SUPPLY AND INSTALLATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY O NMG10027-P016 420001123900 SUPPLY AND INSTALLATION OF HEAVY-DUTY HERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG10027-P017 4200011239100 SUPPLY AND INSTALLATION OF RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200
NMG10027-P010 4200011238400 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO SCRUB NMG10027-P011 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P012 4200011238600 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE NMG10027-P014 4200011238800 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY O NMG10027-P015 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY HERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG10027-P014 420001123900 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P014 420001123900 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTERIAL & ANTIFIER RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 420001123900 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4
NMG10027-P011 4200011238500 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P012 4200011238600 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATON NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE NMG10027-P014 4200011238800 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY O NMG10027-P015 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG10027-P016 420001123900 SUPPLY AND INSTALLAIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTERIAL & ANTIFIER RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239200 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P010 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P010 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P010 4200011239400 SUPPLY AND INSTALL CO
NMG10027-P012 420011238600 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO SCRUB STATO NMG1002 NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE NMG1002 NMG10027-P014 4200011238800 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY ON NMG1002 NMG10027-P015 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG1002 NMG10027-P016 420001123900 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INS
NMG10027-P013 4200011238700 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE NMG1002 NMG10027-P014 4200011238800 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY ON NMG1002 NMG1
NMG10027-P014 420011238800 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY O NMG1002 NMG10027-P015 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG1002 NMG10027-P016 420001123900 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P015 4200011238900 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F NMG10027-P016 4200011239000 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P016 4200011239000 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO SCRUB STATON SINGLE BAY NMG10027-P017 4200011239100 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P017 4200011239100 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P018 4200011239200 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P019 4200011239300 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P019 ADDITIONAL STATEMENT OF THE PROPERTY OF
NMG10027-P020 4200011239400 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG1002
·
NMG10027-P021 4200011239500 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P021 4200011239500 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P022 4200011239600 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P023 4200011239700 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREA NMG1002
NMG10027-P024 4200011239800 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P025 4200011239900 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P026 4200011240000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P027 4200011240100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P028 4200011240200 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P029 4200011240300 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P030 4200011240400 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P031 4200011240500 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P032 4200011240600 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P033 4200011240700 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P034 4200011240800 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P035 4200011240900 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P036 4200011241000 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW NMG1002
NMG10027-P037 4200011241100 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P038 4200011241200 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P039 4200011241300 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RJ45. AS PER HOSPITAL STANDARDS. RELATED TO SCRUB STATON SINGLE BAY NMG1002
NMG10027-P040 4200011241400 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO SCRUB NMG1002
NMG10027-P041 4200011241500 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO SCRUB STAT NMG1002
NMG10027-P042 4200011241600 RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPES, NMG1002
NMG10027-P043 4200011241700 FIRE WATER LINE RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P044 4200011241800 SPRINGKLER HEADS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P044 4200011241800 SPRINGKLER HEADS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P045 4200011241900 DRAIN PIPES RELATED TO SCRUB STATON SINGLE BAY NMG10027-P045 4200011241900 DRAIN PIPES RELATED TO SCRUB STATON SINGLE BAY NMG10027-P045 4200011241900 DRAIN PIPES RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P046 4200011242000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P046 4200011242000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P046 4200011242000 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P047 4200011242100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY NMG10027-P047 4200011242100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY NMG10027-P047 4200011242100 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P047 4200011242200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEFA FILTER) RELATED TO SCROB STATON SINGLE BAY NMG10027-P048 4200011242200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO SCROB STATON SINGLE BAY NMG10027-P048 4200011242200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO SCROB STATON SINGLE BAY NMG10027-P048 4200011242200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO SCROB STATON SINGLE BAY NMG10027-P048 4200011242200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO SCROB STATON SINGLE BAY
NMG10027-P048 4200011242200 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY RETORN). (WITHOUT HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY NMG10027-P049 4200011242300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY NMG10027-P049 4200011242300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY NMG10027-P049 4200011242300 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P050 4200011242400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY NMG10027-P050 4200011242400 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P051 4200011242500 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO SCRUB \$NMG1002
NMG10027-P052 4200011242600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P052 4200011242600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO SCRUB STATON SINGLE BAY NMG10027-P052 4200011242600 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P053 4200011242700 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO SCRUB ST NMG1002
NMG10027-P054 4200011242800 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FOI NMG1002
NMG10027-P055 4200011242900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P055 4200011242900 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-P056 4200011243000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY NMG10027-P056 4200011243000 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO SCRUB STATON SINGLE BAY
NMG10027-W3 4200011658500 3rd YEAR WARRANTY COST RELATED TOSCRUB STATON SINGLE BAY NMG1002
NMG10027-W4 4200011658600 4th YEAR WARRANTY COST RELATED TOSCRUB STATON SINGLE BAY NMG1002
NMG10027-W5 4200011658700 Sth YEAR WARRANTY COST RELATED TOSCRUB STATON SINGLE BAY

NMC10027	43404C4004900 ISCRIUD STATON SINCUS DAV	NIN4C10027
NMG10027	4219161001800 SCRUB STATON SINGLE BAY	NMG10027
NMG10028-D1	400000054300 MAIN AREA DELIVERY COST RELATED TO STRETCHER MORTUARY	NMG10028
NMG10028-D2 NMG10028-W3	400000054400 REMOTE AREA DELIVERY COST RELATED TO STRETCHER MORTUARY 4200011658900 3rd YEAR WARRANTY COST RELATED TO STRETCHER MORTUARY	NMG10028
NMG10028-W3	4200011659000 3th YEAR WARRANTY COST RELATED TOSTRETCHER MORTUARY	NMG10028 NMG10028
NMG10028-W4		
NMG10028-W5	4200011659100 5th YEAR WARRANTY COST RELATED TOSTRETCHER MORTUARY 4226180000000 STRETCHER MORTUARY	NMG10028
NMG10028	4226170300100 WASTE MIXER RELATED TO TABLE AUTOPSY	NMG10028
NMG10031-001	4226170300200 FILTERS KIT RELATED TO TABLE AUTOPSY	NMG10031 NMG10031
NMG10031-002	4226170300300 ORGAN TABLE COMPATIBLE RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-003	400000054300 MAIN AREA DELIVERY COST RELATED TO TABLE AUTOPSY	
NMG10031-D1		NMG10031
	400000054400 REMOTE AREA DELIVERY COST RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P001	4200011243100 PREPARING THE SITE TO START WORK BY REMOVING, DISMANTLING, CRACKING AND MOVING ALL THAT IS NEEDED TO DISMANTLE WALLS, FLOORS AND CEILINGS FROM INSIDE THE BUILD	
NMG10031-P002	4200011243200 SUPPLY AND INSTALLATION OF STAINLESS STEEL DRAINGE GRILL WIDTH NOT LESS THAN 12 CM RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P003	4200011243300 PROVISION OF PAINT WALLS WITH PRIMER AND FILLERS AND THREE-COATED WASHABLE HYGIENIC PAINTS. PAINTING MATERIALS AND COLOR TO MATCH EXISTING WALL PAINT OR AS PER	
NMG10031-P004	4200011243400 SUPPLY AND INSTALLATION OF CABLE TRAY RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P005	4200011243500 SUPPLY AND INSTALLATION OF HARDWARE ACCESORIES FOR THE DOORS, COMPLETE WITH HINGES, DOOR LOCKS, DOOR CLOSER, DOOR STOPPER. ETC RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P006	4200011243600 CLEARING AND CLEANING OF ANY OBSTRUCTIONS WITHIN THE PROPOSED WORK AREA. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P007	4200011243700 DISMANTLING OF THE EXISTING MEDICAL EQUIPMENT. WORKS INCLUDE, WOOD CRATING AND TRANSPORT TO HOSPITAL'S WAREHOUSE. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P008	4200011243800 REMOVAL, DEMOLITION OF EXISTING WALLS, TO ACCOMMODATE THE NEW MEDICAL EQUIPMENT. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P009	4200011243900 REMOVAL AND REINSTATEMENT OF WALLS (EXTERNAL/ INTERNAL)/ WINDOW/ DOOR, AFFECTED TO CLEAR THE MEDICAL EQUIPMENT DELIVERY ROUTE. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P010	4200011244000 CONSTRUCTION OF MASONRY WALLS WITH FIRE RATED AND MOISTURE RESISTANT, FIBER OR ROCKWOOL SOUND-PROOF INSULATION, AS REQUIRED IN THE DRAWING. RELATED TO TABLE	
NMG10031-P011	4200011244100 SUPPLY AND INSTALLATION OF MOTORIZED SWING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL RELATED TO TABLE AUTOPS	
NMG10031-P012	4200011244200 SUPPLY AND INSTALLATION OF MOTORIZED SLIDING DOOR, INCLUDING OVERHEAD MECHANISM, SIZE AS PER DRAWING. MATERIAL AS APPROVED BY HOSPITAL. RELATED TO TABLE AUTOP	
NMG10031-P013	4200011244300 SUPPLY AND INSTALLATION OF SINGLE SWING DOORS, INCLUDING HEAVY DUTY HARDWARE HINGES AND DOOR HANDLES ON HOLLOW METAL DOOR FRAME. SIZE AS PER DRAWING. MATE	
NMG10031-P014 NMG10031-P015	4200011244400 SUPPLY AND APPLICATION OF WASHABLE HYGIENIC SEMI-GLOSS WALL ANTIBACTERIAL PAINT, INCLUDING SURFACE PREPARATION WITH A MINIMUM OF THREE COATS AND PREFERABLY C 4200011244500 SUPPLY AND INSTALLATION OF HEAVY-DUTY THERMAK WELDED ROLLED TYPE ANTI-STATIC VINYL FLOOR COVERING WITH BORDER WELDED SEAMS EXTENDED TO THE WALL WITH COVER F	
NMG10031-P016	4200011244600 SUPPLY AND INSTALL AIR LOCK DOOR SYSTEM RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P017 NMG10031-P018	4200011244700 SUPPLY AND INSTALL ELECRICAL HUNG DOOR OR SLIDING DOORS RELATED TO TABLE AUTOPSY	NMG10031 NMG10031
NMG10031-P018	4200011244800 SUPPLY AND INSTALL SINGLE WOODEN DOOR ANTIBACTRIAL & ANTIFIRE RELATED TO TABLE AUTOPSY 4200011244900 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P019	4200011245000 SUPPLY AND INSTALL STAINLESS STEEL WALL PANELING AROUND EQUIPMENTS RELATED TO TABLE AUTOPSY 4200011245000 SUPPLY AND INSTALL COMPRESSOR VIBRATION RESISTANCE BASE AS PER HOSPITAL REQUIREMENTS RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P020	4200011245100 SUPPLY AND INSTALL COMPRESSOR SHADING COVER AS PER HOSPITAL REQUIREMENTS RELATED TO TABLE AUTOPSY	NMG10031 NMG10031
NMG10031-P021	4200011245100 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO TABLE AUTOPSY	NMG10031 NMG10031
NMG10031-P022	4200011245200 SUPPLY AND INSTALL COMPRESSOR CAGE AS PER HOSPITAL REQUIREMENTS RELATED TO TABLE AUTOPSY 4200011245200 SUPPLY AND INSTALL PDB (POWER DISTRIBUTION BOARD) PER FINAL/ELECTRICAL DRAWINGS FOR ENTIRE SUITE, COMPLETE WITH THE REQUIRED CABLES AND ACCESSORIES, CIRCUIT BREA	
NMG10031-P023	4200011245300 INSTALLATION OF ANTI BACTERIAL LED LIGHTING FIXTURES (60 CM X 60 CM) WITH SWITCHES AND WIRING DEVICES AS PER HOSPITAL STANDARDS. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P025	4200011245500 INSTALLATION OF POWER SOCKET COMPLETE WITH ALL ELECTRICAL CONDUITS, WIRING, BOXES, ETC RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P026	4200011245600 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 35 MM2 CABLE WIRES RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P027	4200011245700 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 50 MM2 CABLE WIRES RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P028	4200011245800 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 70 MM2 CABLE WIRES RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P029	4200011245900 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 90 MM2 CABLE WIRES RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P030	4200011246000 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 120 MM2 CABLE WIRES RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P031	4200011246100 SUPPLY AND INSTALLATION OF THE REQUIRED XLPE/PVC POWER CABLES; 3 PHASE, 380V, 185 MM2 CABLE WIRES RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P032	4200011246200 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 200 AMP RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P033	4200011246300 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 250 AMP RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P034	4200011246400 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 300 AMP RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P035	4200011246500 SUPPLY AND INSTALLATION OF MAIN DISCONNECT SWITCH IN THE EQUIPMENT ROOM. 400 AMP RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P036	4200011246600 SUPPLY AND INSTALLATION OF MAINS DISTRIBUTION PANEL BOARD (MDP), COMPLETE WITH THE REQUIRED CONNECTION CABLES AND ACCESSORIES, GFI, CIRCUIT BREAKER, SLOW BLOW 4200011246700 SUPPLY AND INSTALLATION OF CABLE TRAY INSIDE THE FALSE CEILING (KW). RELATED TO TABLE AUTOPSY	
NMG10031-P037		NMG10031
NMG10031-P038 NMG10031-P039	4200011246800 SUPPLY AND INSTALLATION OF POWER OUTLET COMPLETE WITH CABLES. AS PER HOSPITAL STANDARDS. RELATED TO TABLE AUTOPSY 4200011246900 SUPPLY AND INSTALLATION OF TELEPHONE OUTLET COMPLETE WITH RJ45. AS PER HOSPITAL STANDARDS. RELATED TO TABLE AUTOPSY	NMG10031 NMG10031
NMG10031-P040	4200011247000 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITHOUT HEPA FILTER) RELATED TO TABLE A	
NMG10031-P041	4200011247100 SUPPLY INSTALLATION OF THE A/C DIFFUSERS COMPLETE WITH PLENUM BOX, FLEXIBLE, COLAR AND AS PER HVAC DRAWING AND LOCATION. (WITH HEPA FILTER) RELATED TO TABLE AUTO	
NMG10031-P042 NMG10031-P043	4200011247200 RELOCATION/ DIVERSION OF ANY EXISTING UTILITY PIPES WITHIN THE CEILING VOID AREA DIRECTLY ABOVE THE EQUIPMENT, SUCH AS FIRE WATER LINE, SPRINGKLER HEADS, DRAIN PIPES,	
INIVIG 10031-P043	4200011247300 FIRE WATER LINE RELATED TO TABLE AUTOPSY	NMG10031

NMG10031-P044	4200011247400 SPRINGKLER HEADS RELATED TO TABLE AUTOPSY	NMG10031
-	4200011247500 DRAIN PIPES RELATED TO TABLE AUTOPSY	NMG10031
	4200011247600 SUPPLY AND INSTALLATION OF A/C DUCTS. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P047	4200011247700 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITH HEPA FILTER) RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P048	4200011247800 SUPPLY AND INSTALLATION OF A/C DIFFUSERS (SUPPLY/ RETURN). (WITHOUT HEPA FILTER) RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P049	4200011247900 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITHOUT HEPA FILTER) RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P050	4200011248000 SUPPLY AND INSTALLATION OF A/C GRILLS. (WITH HEPA FILTER) RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P051	4200011248100 CONNECTIVITY OF A/C UNIT TO EXISTING HOSPITAL BMS FOR MONITORING AND CONTROL. WORKS NOT LIMITED TO WIRING, ALL ACCESSORIES AND PROGRAMMING. RELATED TO TABLE /	NMG10031
NMG10031-P052	4200011248200 SUPPLY OF PORTABLE CO2 FIRE EXTINGUISHER. RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P053	4200011248300 SUPPLY AND INSTALLATION OF WATER SUPPLY AND FLOOR DRAINAGE SYSTEM WHICH INCLUDE PIPING WORKS, TAPPED TO THE NEAREST BUILDING UTILITY SYSTEM. RELATED TO TABLE A	NMG10031
NMG10031-P054	4200011248400 BE SUPPLYING AND INSTALLING PRESSURE MONITORS THAT CAN PROVIDE FOUR READINGS (PRESSURE, TEMPERATURE, HUMIDITY AND AIR CHAMGE PER HOUR) AND STORE THE DATA FO	NMG10031
NMG10031-P055	4200011248500 SUPPLY AND INSTALL CHILER SYSTEM AS PER HOSPITAL REQUIREMENTS RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-P056	4200011248600 SUPPLY HOT & COLD WATER FOR MORGUE EQUIPMENTS RELATED TO TABLE AUTOPSY	NMG10031
NMG10031-W3	4200011659300 3rd YEAR WARRANTY COST RELATED TOTABLE AUTOPSY	NMG10031
NMG10031-W4	4200011659400 4th YEAR WARRANTY COST RELATED TOTABLE AUTOPSY	NMG10031
NMG10031-W5	4200011659500 Sth YEAR WARRANTY COST RELATED TOTABLE AUTOPSY	NMG10031
NMG10031	4226170300000 TABLE AUTOPSY	NMG10031
NMG10033-001	4226180000100 BOWL TO COLLECT WASTE RELATED TO TROLLEY VIEWING	NMG10033
NMG10033-002	4226160400000 HEAD REST RELATED TO TROLLEY VIEWING	NMG10033
NMG10033-D1	400000054300 MAIN AREA DELIVERY COST RELATED TO TROLLEY VIEWING	NMG10033
NMG10033-D2	400000054400 REMOTE AREA DELIVERY COST RELATED TO TROLLEY VIEWING	NMG10033
NMG10033-W3	4200011659700 3rd YEAR WARRANTY COST RELATED TOTROLLEY VIEWING	NMG10033
NMG10033-W4	4200011659800 4th YEAR WARRANTY COST RELATED TOTROLLEY VIEWING	NMG10033
NMG10033-W5	4200011659900 5th YEAR WARRANTY COST RELATED TOTROLLEY VIEWING	NMG10033
NMG10033	4219240403300 TROLLEY VIEWING	NMG10033
NMG10034-D1	400000054300 MAIN AREA DELIVERY COST RELATED TO TROLLEY WASHING BODY	NMG10034
NMG10034-D2	400000054400 REMOTE AREA DELIVERY COST RELATED TO TROLLEY WASHING BODY	NMG10034
NMG10034-W3	4200011660100 3rd YEAR WARRANTY COST RELATED TOTROLLEY WASHING BODY	NMG10034
NMG10034-W4	4200011660200 4th YEAR WARRANTY COST RELATED TOTROLLEY WASHING BODY	NMG10034
NMG10034-W5	4200011660300 5th YEAR WARRANTY COST RELATED TOTROLLEY WASHING BODY	NMG10034
NMG10034	4219240403400 TROLLEY WASHING BODY	NMG10034
NMG10038-D1	400000054300 MAIN AREA DELIVERY COST RELATED TO LIFTER BODY MANUAL	NMG10038
NMG10038-D2	400000054400 REMOTE AREA DELIVERY COST RELATED TO LIFTER BODY MANUAL	NMG10038
NMG10038-W3	4200011660500 3rd YEAR WARRANTY COST RELATED TOLIFTER BODY MANUAL	NMG10038
NMG10038-W4	4200011660600 4th YEAR WARRANTY COST RELATED TOLIFTER BODY MANUAL	NMG10038
NMG10038-W5	4200011660700 Sth YEAR WARRANTY COST RELATED TOLIFTER BODY MANUAL	NMG10038
NMG10038	4226210000100 LIFTER BODY MANUAL	NMG10038